



Description

The NC-103 Pull Station is the perfect accessory to the MiCare advanced nurse call solution. A fully standalone wireless device that can be placed anywhere in a facility, the NC-103 has two standard user serviceable AA batteries for long battery life.

The stylish design follows the same lines as the NC-2000 Patient Unit and boasts 3 buttons, pull cord and external input from the rear. An easy to identify main emergency button, a germ wipe-able pull cord that will never tangle and always hangs straight. Two additional buttons for staff assist and room maintenance request. Two terminal blocks on the rear of the unit provide an additional input for door contact or flood detectors in bathroom areas.

It easily mounts in a single gang PVC box and comes with a locking bolt to secure the device after installation. Attach the mounting plate first and snap in place.

Features

- Sleek ergonomic design
- 3 Buttons, Pull Cord and additional input provide many options for configuration
- Long battery life with two standard AA batteries
- Batteries are easily changed by maintenance staff
- Low battery indicator which routes to special reports for maintenance staff and their own electronic checklist of completed maintenance
- Germ Wipe-able medical grade plastics
- Can be placed anywhere in the facility
- Fully supervised heartbeats to Central ensuring 24/7 operation
- LED indicator
- Audible indicator
- Cancel at source functionality

Technical Specifications

Standards:	UL 1069, 2560, CSA 22.2 #205, FCC Part 15, Class "B", CS-03
Power Options:	Input Power: <ul style="list-style-type: none"> • (3.0V) 2 x 1.5V AA Batteries • Nominal Current 35mA
Wireless Characteristics:	Frequency: 2.4Ghz Network type: 802.15.4 wireless mesh network Range: 75ft
Physical Characteristics:	Faceplate: 221R Lexan Impact Strength: 260 ft-lbs/in2 Color: Off White Weight: 4oz (113g) Dimensions: 5.1"H x 3.5" W x 0.8" D Standard 1-gang, 2.5" Shallow PVC Back Box Required Operating Temperature: 0° C to +50° C
Input Operating Specifications:	Wet contact pair, supplied voltage is max 3.6V Nominal current of 1.4mA (external loop closed) and an absolute maximum of 5mA. Follow manufacturer's specifications for any attached equipment or wiring.

Ordering Information

Model	Description
NC-103	Pull Station

NOT TO BE USED FOR INSTALLATION PURPOSES.



Canada
 25 Interchange Way
 Vaughan, Ontario L4K 5W3
 Telephone: (905) 660-4655
 Fax: (905) 660-4113

U.S.A.
 4575 Witmer Industrial Estates
 Niagara Falls, NY 14305
 Toll Free: (888) 660-4655
 Fax Toll Free: (888) 660-4113

Distributed by: